

## Evaluation Board for the AD7711, 24-Bit Sigma-Delta A/D Converter

by Mike Byrne and Albert O'Grady

### INTRODUCTION

This application note describes the evaluation board for the AD7711 24-bit signal conditioning A/D converter. This converter utilizes sigma-delta techniques to offer 16-bit accuracy (0.0015% FSR) and no missing codes to 24 bits. It contains a two-channel programmable gain front end, a programmable low-pass digital filter and a bidirectional serial interface. One of the analog inputs and the reference input are differential, and the part can accept inputs as low as 20 mV full scale. It features a self-calibration mode that removes internal offset and gain errors, a system calibration mode that removes external circuit offset and gain errors, and a background calibration mode that automatically removes internal offset and gain errors. In addition, its flexible serial interface allows the AD7711 to connect directly to digital signal processors (ADSP-2101, TMS320C25, etc.) and microcontrollers (8XC51, 68HC11, etc.). Full data on the AD7711 is provided in the AD7711 data sheet available from Analog Devices and should be

consulted in conjunction with this application note when using the evaluation board. On-board components include an AD680, +2.5 V reference (the part's own +2.5 V reference can also be used), a 10 MHz crystal (an external clock source can also be used), and bus buffers to facilitate the AD7711's two serial interface modes.

The basic evaluation board schematic is shown in Figure 1 while a full circuit diagram for the board is shown in Figure 3. The component silkscreen, component side layout and solder side layout for the evaluation board are given in Figures 4, 5 and 6 respectively.

### LINK OPTIONS

The AD7711 Evaluation Board has several link options available to facilitate a variety of operating conditions and these are summarized in Table I. The position of the link options should be chosen before any power is applied to the evaluation board.




Figure 1. Evaluation Board Basic Schematic Diagram

Table I. Link Options

Link No.	Function
LK1	<p>This option selects the serial interface mode for the AD7711. It is a double link (both links must be moved together for correct operation) which sets the AD7711 MODE pin and configures the bidirectional buffer on the serial clock.</p> <p>With both links in Position A, the MODE input is low putting the device into its external clocking mode; the AD7711 SCLK pin is configured as an input, and the serial clock buffer (74HC125) is configured as an input.</p> <p>With both links in Position B, the MODE input is high putting the device into its self-clocking mode; the AD7711 SCLK pin is configured as an output, and the serial clock buffer (74HC125) is also configured as an output.</p>
LK2	<p>This option selects the master clock source for the AD7711. The master clock source comes from the on-board crystal or from an external clock source via SKT6. It is a double link (both links must be moved together for correct operation).</p> <p>With both links in Position A, the external clock is selected. The external clock for the AD7711 must be applied to SKT6 which is routed to the MCLK IN pin of the AD7711.</p> <p>With both links in Position B, the on-board 10 MHz crystal (XTAL 1) provides the master clock for the device. With the links in this position, the crystal is connected between the MCLK IN and MCLK OUT pins of the AD7711.</p>
LK3	<p>This link is used to short out the resistor in the reference decoupling scheme. With this link option in place the reference decoupling resistor, R4, is shorted out. When the AD680 reference is used as the reference source for the device, this link option should not be in place. When using the AD7711's own internal reference or when using an external reference which does not require any decoupling resistor, this link should be in place.</p>
LK4	<p>This option is used to select the reference source for the AD7711's REF IN(+) pin.</p> <p>With this link in Position A, the AD7711's own on-chip reference is used as the reference source for the AD7711 and REF OUT is connected directly to the REF IN(+) input.</p> <p>With this link in Position B, the AD680 +2.5 V reference is used as the reference source for the AD7711 and the AD680's V<sub>OUT</sub> pin is connected, via the decoupling components, to the REF IN(+) input.</p> <p>With this link in Position C, an external reference, applied to SKT7, can be used as the reference source for the AD7711 and SKT7 is connected, via the decoupling components (see LK3), to the REF IN(+) input.</p>
LK5	<p>This option is used to select the voltage source for the AD7711's REF IN(-) pin.</p> <p>With this link in Position A, the REF IN(-) pin of the AD7711 is connected directly to AGND. If the AD7711's own reference is used as the reference source for REF IN(+), REF IN(-) should be connected to AGND.</p> <p>With this link in Position B, an external voltage, applied to SKT8, can be used as the reference source for the AD7711's REF IN(-) input and SKT8 is connected directly to the REF IN(-) input. If an external differential voltage is required for the AD7711 it should be applied between SKT7 and SKT8 with LK4 and LK5 set appropriately.</p>
LK6	<p>This option is used to select the voltage source for the AD7711's V<sub>BIAS</sub> pin.</p> <p>With this link in Position A, the AD7711's own on-chip reference is used as the voltage source for the V<sub>BIAS</sub> input and REF OUT is connected directly to the V<sub>BIAS</sub> input.</p> <p>With this link in Position B, the V<sub>BIAS</sub> input is connected directly to AGND.</p> <p>With this link in Position C, an external voltage, applied to SKT5, can be used as the voltage source for the V<sub>BIAS</sub> input and SKT9 is connected directly to the V<sub>BIAS</sub> input.</p>
LK7-LK9	<p>These links are in series with the AIN1(+), AIN1(-), and AIN2 analog inputs respectively.</p> <p>With link LK7 in place, SKT2 is connected directly to the AIN1(+) input. With LK8 in place, SKT3 is connected directly to the AIN1(-) input. With link LK9 in place, SKT4 is connected directly to the AIN2 input. These links may be removed so that the input signals at SKT2 to SKT4 can be connected to the component grid for signal conditioning before being applied to the analog inputs of the AD7711.</p>
LK10	<p>This link is used to short out the Schottky diode, D1, on the AV<sub>DD</sub> supply when the device is operated with AV<sub>DD</sub> and DV<sub>DD</sub> driven from the same supply. When AV<sub>DD</sub> and DV<sub>DD</sub> are driven from different supplies, and the short-circuit current capability of the DV<sub>DD</sub> supply is too large for the Schottky diode, D2, then link option LK10 should be removed placing diode D1 in series with the AV<sub>DD</sub> line.</p>

## POWER SUPPLIES AND GROUNDING

The AD7711 has three power supply inputs:  $AV_{DD}$ ,  $DV_{DD}$  and  $V_{SS}$ . The  $AV_{DD}$  range is from +5 V to +10 V,  $DV_{DD}$  is +5 V and  $V_{SS}$  is 0 V or -5 V. There is an additional requirement that  $AV_{DD}-V_{SS}$  must be less than 10.5 V. The evaluation board caters for all options by providing  $AV_{DD}$ ,  $DV_{DD}$  and  $V_{SS}$  inputs. If  $AV_{DD} = +5$  V and  $DV_{DD} = +5$  V, it is recommended that  $AV_{DD}$  and  $DV_{DD}$  are driven from the same +5 V supply to avoid any power sequencing issues. In applications where two separate supply voltages are required for  $AV_{DD}$  and  $DV_{DD}$ , the 10  $\Omega$  resistor, R9, and the SD103C Schottky diode, D2, provide protection for the AD7711. In applications where the short-circuit current capability of the  $DV_{DD}$  supply is too large for the SD103C, link option LK10 should be removed placing diode D1 in series with the  $AV_{DD}$  line.

All supplies are decoupled to ground with 10  $\mu$ F tantalum and 0.1  $\mu$ F ceramic disc capacitors. The  $AV_{DD}$  and  $V_{SS}$  supplies are decoupled to the AGND plane while the  $DV_{DD}$  supply is decoupled to the DGND plane. Power for the digital section of the board and the  $DV_{DD}$  pin of the AD7711 is routed separately from the power for the analog section and  $AV_{DD}$  on the AD7711. For systems using a single +5 V supply for  $AV_{DD}$  and  $DV_{DD}$ , the  $AV_{DD}$  and  $DV_{DD}$  input terminals should be connected together.

The evaluation board uses extensive ground planing to minimize any high frequency noise interference from the on-board clocks or any other sources. Once again, the ground planing for the analog section is kept separate from that for the digital section, and they are joined only at the AD7711 AGND and DGND pins.

## ANALOG INPUT SECTION

The AD7711 provides two analog input channels, one differential and one single-ended. The analog inputs to the evaluation board are applied to the subminiature BNC connectors labelled SKT2, SKT3 and SKT4. With links LK7 through LK9 in place these BNC connector inputs are routed to the AIN1(+), AIN1(-) and AIN2 inputs of the AD7711. These links can be removed to allow the input signals at SKT2 to SKT4 to be routed to the component grid if input signal conditioning is required. There is also the option of placing R, C antialiasing filters on each analog input line (labelled R1 through R3 and C11 through C13 on the silkscreen).

## REFERENCE INPUT

The reference input for the AD7711 is a differential input with the reference voltage,  $V_{REF}$ , equal to the difference between the REF IN(+) pin and the REF IN(-) pin. The evaluation board allows the reference to be generated from three sources using link options, LK4 and LK5. Table II summarizes the reference link options. There is an additional link, LK3, also used with the reference. When in place, LK3 shorts out the reference decoupling resistor, R4. This decoupling resistor is required when using the AD680's  $V_{OUT}$  pin as the reference source for REF IN (+). It may also be required by external reference sources. The

REF IN(+) input is decoupled to AGND via C14, a 0.1  $\mu$ F ceramic disc capacitor.

The first of the reference options is to derive the reference voltage from the AD7711's own internal reference. R4 can be shorted out using LK3 when REF OUT is driving REF IN(+). REF OUT provides a single-ended, +2.5 V reference which is referred to AGND. Therefore, when REF OUT is connected to REF IN(+), the REF IN(-) input should be connected to AGND via LK5. The second option is to use the AD680's  $V_{OUT}$  pin as the reference source for REF IN(+). In this case, LK3 should be removed. Once again, the AD680 provides a single-ended, +2.5 V reference which is referred to AGND. Therefore, when the AD680  $V_{OUT}$  pin is connected to REF IN(+), the REF IN(-) input should be connected to AGND via LK5.

The third option is to use an external voltage as the reference source for the part. This external reference voltage is applied to the evaluation board at SKT7. This external reference voltage can be single-ended (in which case REF IN(-) is connected to AGND via LK5) or it can be a differential voltage applied between SKT7 and SKT8. LK5 can route the voltage on SKT8 directly to the REF IN(-) input of the AD7711.

Table II. Reference Link Options

LK4 Position	LK5 Position	REF IN(+) Voltage	REF IN(-) Voltage
A	A	REF OUT	AGND
B	A	AD680 $V_{OUT}$	AGND
C	A	SKT7 I/P	AGND
C	B	SKT7 I/P	SKT8 I/P

## $V_{BIAS}$ INPUT

The  $V_{BIAS}$  input for the AD7711 provides the return path for most of the currents flowing in the analog modulator. It provides the bias point for much of the analog circuitry in the modulator. The limits on the  $V_{BIAS}$  voltage depend on the reference voltage and on the  $AV_{DD}$  and  $V_{SS}$  voltages. Consult the data sheet to ensure correct biasing. The evaluation board allows the  $V_{BIAS}$  voltage to be generated from three sources using link option, LK6. Table III summarizes the  $V_{BIAS}$  link options.

The first of the options connects the  $V_{BIAS}$  input directly to the REF OUT pin of the AD7711. This connection point can be used for either single supply systems ( $V_{SS} = 0$  V) or dual supply systems ( $V_{SS} = -5$  V) with a  $V_{REF}$  of +2.5 V. The second option connects the  $V_{BIAS}$  input directly to AGND. This connection point can be used for dual supply systems with a  $V_{REF}$  of +2.5 V. The third option connects the  $V_{BIAS}$  input to SKT5 to allow an external bias voltage to be applied to the AD7711. Note, the  $V_{BIAS}$  input should be driven from a low impedance point.

**Table III.  $V_{BIAS}$  Link Options**

LK6 Position	$V_{BIAS}$ Input
A	REF OUT
B	AGND
C	SKT5 I/P

The  $V_{BIAS}$  voltage does have an effect on the  $AV_{DD}$  power supply rejection performance of the AD7711. If the  $V_{BIAS}$  voltage tracks the  $AV_{DD}$  supply, it improves the power supply rejection from the  $AV_{DD}$  supply line from 80 dB to 95 dB typ. Using an external Zener diode, connected between the  $AV_{DD}$  line and the  $V_{BIAS}$  input, as the source for the  $V_{BIAS}$  voltage gives the improvement in the  $AV_{DD}$  power supply rejection performance.

### CLOCK GENERATION

There are two clock source options for the master clock for the AD7711 selected by link LK2. This is a double link option, and the device will not operate correctly unless both links of LK2 are in the same position. When both links are in Position B, the on-board 10 MHz crystal oscillator is connected between the MCLK IN and MCLK OUT pins of the AD7711, and this generates the master clock for the AD7711. With both links in position A, MCLK OUT is left open-circuit, and MCLK IN is connected directly to SKT6. An external, CMOS-compatible, clock applied to SKT6 will provide the master clock for the AD7711. The frequency of this clock must be in the range from 400 kHz to 10 MHz. Note, for  $AV_{DD}$  voltages in excess of +5.25 V, the AD7711 is only specified to operate with a master clock frequency up to 8 MHz.

### AD7711 INTERFACE MODE SELECTION

The AD7711 may be set up for either of two serial interface modes. These are controlled on the evaluation board by link option LK1. This is a double link option and the device will not operate correctly in either of the modes unless both links of LK1 are in the same position. The modes are as follows:

1. The Self-Clocking Mode can be used with processors which allow an external device to clock their serial port including most digital signal processors and microcontrollers such as the 68HC11 and 68HC05. It also allows easy interfacing to serial-parallel conversion circuits, allowing interfacing to 74XX299 Universal Shift registers without any additional decoding. In this mode, the AD7711 acts as the master device in the serial system, and data is clocked out of the part on the falling edge of the internally generated SCLK and is clocked into the device during the high time of the same internally generated clock.

2. The External Clocking Mode, is designed for systems which provide a serial clock output which is synchronized to the serial data output including microcontrollers such as the 8XC51, 68HC11 and 68HC05 and most digital signal processors. In this mode, the AD7711 acts as a slave device in the serial system, and data is clocked out of the part on the falling edge of the externally applied SCLK and is clocked into the device during the high time of the same externally applied clock.


Table IV gives the link option positions for the interface modes. Note, inputs  $A0$ ,  $\overline{SYNC}$ ,  $\overline{TFS}$  and  $\overline{RFS}$  all have 10 k $\Omega$  pull-up resistors to  $DV_{DD}$ .

**Table IV. Interface Mode Selection**

LK1 Position	Interface Mode
A	External-Clocking Mode
B	Self-Clocking Mode

### EVALUATION BOARD INTERFACING

Interfacing to the evaluation board is via a 9-way D-type connector, SKT1. The pinout for this connector is shown in Figure 2, its pin designations are given in Table V, and the pin function descriptions are given in Table VI. The port can be used in both the Self-Clocking Mode and the External Clocking Mode. The port has nine lines which are described below.

**Figure 2. Pin Configuration for SKT1, D-Type Connector****Table V. SKT1 Pin Function**

Pin No.	Mnemonic
1	SCLK
2	$\overline{DRDY}$
3	$\overline{RFS}$
4	$\overline{TFS}$
5	$A0$
6	DGND
7	SDATA
8	$\overline{DV_{DD}}$
9	$\overline{SYNC}$

**Table VI. SKT1 Pin Function Description**

Mnemonic	Description
SCLK	Serial Clock Input/Output. When the device is operating in the External-Clocking Mode, the external serial clock is applied to this terminal and goes via IC3 (74HC125 Quadruple Bus Buffer) to the SCLK pin of the AD7711. When the device is operating in the Self-Clocking Mode, the internal serial clock of the AD7711 is routed to this terminal via IC3. The MODE line controls the direction of buffering on this SCLK line.
$\overline{\text{DRDY}}$	This output is the $\overline{\text{DRDY}}$ signal from the AD7711 buffered via IC4 (74HC4050 Hex Buffer).
$\overline{\text{RFS}}$	This buffered input controls the AD7711 $\overline{\text{RFS}}$ input and in conjunction with $\overline{\text{TFS}}$ controls the direction of the buffering on the SDATA line. This $\overline{\text{RFS}}$ input has a 10 k $\Omega$ pull-up resistor to DV <sub>DD</sub> so that the $\overline{\text{RFS}}$ line will be pulled up to its inactive high state even if the $\overline{\text{RFS}}$ input terminal is unconnected.
$\overline{\text{TFS}}$	This buffered input controls the AD7711 $\overline{\text{TFS}}$ input and in conjunction with $\overline{\text{RFS}}$ controls the direction of the buffering on the SDATA line. This $\overline{\text{TFS}}$ input has a 10 k $\Omega$ pull-up resistor to DV <sub>DD</sub> so that the $\overline{\text{TFS}}$ line will be pulled up to its inactive high state even if the $\overline{\text{TFS}}$ input terminal is unconnected.
A0	This buffered input controls the AD7711 A0 input. This A0 input has a 10 k $\Omega$ pull-up resistor to DV <sub>DD</sub> so that the A0 line will be pulled up to a logic high state even if the A0 input terminal is unconnected.
DGND	Digital Ground. This line is connected to the digital ground plane on the evaluation board. It allows the user to provide the digital supply via the connector along with the other digital signals.
SDATA	Serial Data Input/Output. When writing to the AD7711 ( $\overline{\text{TFS}}$ is low), the external serial data is applied to this terminal and goes via IC3 (74HC125 Quadruple Bus Buffer) to the SDATA pin of the AD7711. When reading from the AD7711 ( $\overline{\text{RFS}}$ is low), the internal serial data output of the AD7711 is routed to this terminal via IC3.
DV <sub>DD</sub>	Digital +5 V Supply. This line is connected to the DV <sub>DD</sub> supply line on the evaluation board. It allows the user to provide the digital supply via the connector along with the other digital signals.

Mnemonic	Description
$\overline{\text{SYNC}}$	This buffered input controls the AD7711 $\overline{\text{SYNC}}$ input. This $\overline{\text{SYNC}}$ input has a 10 k $\Omega$ pull-up resistor to DV <sub>DD</sub> so that the $\overline{\text{SYNC}}$ line will be pulled up to its inactive high state even if the $\overline{\text{SYNC}}$ input terminal is unconnected.

**SOCKETS**

There are nine sockets on the evaluation board. The function of these sockets is outlined in Table VII.

**Table VII. Socket Function**

Socket	Function
SKT1	9-Way D-Type Connector
SKT2–SKT4	Subminiature BNC Sockets for Analog Inputs
SKT5	Subminiature BNC Socket for External V <sub>BIAS</sub>
SKT6	Subminiature BNC Socket for External Master Clock
SKT7	Subminiature BNC Socket for External REF IN(+)
SKT8	Subminiature BNC Socket for External REF IN(–)

**SETUP CONDITIONS**

Care should be taken before applying power and signals to the evaluation board to ensure that all link positions are as per the required operating mode. Figure 4 gives the physical layout of all the links on the board in order to ease setup.




Figure 3. AD7711 Evaluation Board Circuit

## REQUIRED COMPONENT LIST

### Integrated Circuits

IC1	AD7711
IC2	AD680 Voltage Reference
IC3	74HC125 Quad Bus Buffers with Three State Outputs
IC4	74HC4050 Hex Buffer

### Capacitors

C1, C3, C5, C7	10 $\mu$ F Capacitors
C2, C4, C6, C8, C9, C10, C14	0.1 $\mu$ F Capacitors

### Resistors

R4	200 $\Omega$ Resistor
R5, R6, R7, R8	10 k $\Omega$ Pull-Up Resistors
R9	10 $\Omega$ Resistor

### Optional Filter Components

R1 C11, R2 C12, R3 C13	Can Be Used on Analog Inputs for Filtering.
------------------------	---

### Crystals

XTAL 1	10 MHz Crystal
--------	----------------

### Links

LK1, LK2, LK3, LK4, LK5, LK6, LK7, LK8, LK9, LK10	Shorting Plugs
---	----------------

### Sockets

SKT1	9-Way D Type Connector
SKT2 to SKT8	Subminiature BNC Sockets

### Diodes

D1	HP5082-2810, Schottky Diode
D2	SD103C, Schottky Diode




Figure 4. AD7711 Evaluation Board Silkscreen




Figure 5. PCB Component Side Layout for Figure 3




Figure 6. PCB Solder Side Layout for Figure 3